



Mask/Device Compatibility List

English

This document specifies the masks compatible with ResMed's positive airway pressure devices and the mask setting, if one is required. For full details regarding the correct use of these devices, please refer to the specific User Guides.

Note: Not all products are available in all regions.

Full Face Masks 1

Device	AcuCare F1-4	AirFit [™] F40 ⁴	AirTouch™ F30i ^{1,4}	AirFit™ F30i¹ / AirFit™ F30i Non Magnetic¹	AirFit [™] F30	AirTouch™ F20 / AirTouch™ F20 for Her	AirFit™ F20 /AirFit™ F20 for Her/ AirFit™ F20 Non Magnetic	AirFit™ F10 / AirFit™ F10 for Her
AirCurve™ 11 VAuto / AirCurve™ 11 S / AirCurve™ 11 ST	Х	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 11 ASV PaceWave	Χ	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirSense [™] 11	N/V	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirMini™	Χ	Χ	Χ	Χ	~	~	✓	X
AirStart™ 10 / AirSense™ 10	N/V	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 VAuto / AirCurve™ 10 S /AirCurve™ 10 ST	N/V	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 CS Pacewave	N/V	Pillows	Full Face	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 CS-A Pacewave	N/V	N/V	N/V	N/V	Full Face	Full Face	Full Face	Full Face
Lumis™	N/V	Pillows ³	Full Face ³	Full Face ³	Full Face ³	Full Face	Full Face	Full Face
S9 AutoSet™ / S9 Elite™	N/V	Pillows	Full Face	Full Face	N/V	Full Face	Full Face	Full Face
S9 Escape™	N/V	Pillows	Full Face	Full Face	N/V	Full Face	Full Face	Full Face
AutoSet CS™ 2	N/V	N/V	N/V	N/V	N/V	N/V	N/V	FULL FACE
AutoSet CS™ (S9)	N/V	Pillows	N/V	N/V	N/V	Full Face	Full Face	Full Face
AutoSet CS™ (S9 PaceWave)	N/V	Pillows	N/V	N/V	N/V	Full Face	Full Face	Full Face
AutoSet CS-A™ (S9 PaceWave)	N/V	N/V	N/V	N/V	N/V	Full Face	Full Face	Full Face
C-Series Tango™	N/V	N/V	N/V	N/V	N/V	N/V	N/V	No setting required. Consult your clinician before changing to a different mask type.
S7™ Lightweight	N/V	N/V	N/V	N/V	N/V	N/V	N/V	No setting required. Consult your clinician before changing to a different mask type.
S7 Elite™ / S7 AutoSet Spirit	N/V	N/V	N/V	N/V	N/V	N/V	N/V	MIR FULL
S7 AutoSet Respond™	N/V	N/V	N/V	N/V	N/V	N/V	N/V	MIRAGE
S8™ Auto 25	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
S8 AutoSet Spirit™	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
S8 AutoSet Vantage™	N/V	N/V	N/V	N/V	N/V	MIR FULL	MIR FULL	MIR FULL
S8 Escape™	N/V	N/V	N/V	N/V	N/V	N/V	N/V	Requires pressure calibration at 12 cm H2O. Consult your clinician and check the user guide.

Device	AcuCare F1-4	AirFit™404	AirTouch™ F30	j AirFit™ F30i1 / AirFit™ F30i No Magnetic1		AirTouch™ F20 / AirTouch™ F20 for H	AirFit™ F20 /AirFit™ F Hear Her/ AirFit™ F20 Non Magnetic	F20 AirFit™ F10 / AirFit™ F10 for Her
S8 Escape™ II S8	N/	N/V	N/V	N/V	N/	Requi	irespressurecalibrationat1	12cm H2O. Consult your clinician and check the user guide.
Elite™ / S8 Elite™ II S8	V	N/V	N/V	N/V	V	MIR FULL	MIR FULL	MIR FULL
AutoSet Spirit™ II S8	N/	N/V	N/V	N/V	N/	MIR FULL	MIR FULL	MIR FULL
Compact™ VPAP™ III /	V	N/V	N/V	N/V	V	N/V	N/V	No setting required. Consult your clinician before changing to a different r
VPAPTM III ST VPAPTM	N/	N/V	N/V	N/V	N/	N/V	N/V	MIR FULL
III ST-A VPAPTM IV /	V	N/V	N/V	N/V	V	N/V	N/V	MIR FULL
VPAPTM IV ST VPAPTM	N/	N/V	N/V	N/V	N/	MIR FULL	MIR FULL	MIR FULL
Adapt SV VPAP™ Adapt	V	N/V	N/V	N/V	V	N/V	N/V	FULL FACE
(S9) VPAP™ Adapt (S9	N/	Pillows	Full Face	Full Face	N/	Full Face	Full Face	Full Face
PaceWave) VPAP TM	V	Pillows	N/V	Full Face	V	Full Face	Full Face	Full Face
Malibu VPAP™ Auto 25	N/	N/V	N/V	N/V	N/	N/V	N/V	FULL FACE
Auto 25 (S9) VPAPTM S /	V	N/V	N/V	N/V	V	MIR FULL	MIR FULL	MIR FULL
VPAPTM ST VPAPTM S	N/	Pillows	N/V	Full Face	N/	Full Face	Full Face	Full Face
(S9) / VPAPTM ST (S9)	V	N/V	N/V	N/V	V	MIR FULL	MIR FULL	MIR FULL
VPAPTM ST-A (S9)	N/	Pillows	Full Face	Full Face	N/	Full Face	Full Face	Full Face
VPAP™ ST (S9 iVAPS) /	V	N/V	N/V	N/V	V	Full Face	Full Face	Full Face
VPAPTM ST-A (S9 iVAPS)	N/ V	N/V	N/V	N/V	N/ V	Full Face	Full Face	Full Face
S9 VPAP™ Tx VPAP™ Tx	V N/	Pillows2	Full Face2	Full Face2	V N/	Full Face	Full Face	Full Face
	V	N/V	N/V	N/V	V	N/V	N/V	FULL FACE
Stellar TM	N/	N/V	N/V	N/V	N/	FULL FACE	FULL FACE	FULL FACE
VS Ultra™ / VS Integra™	V	N/V	N/V	N/V	V	N/V	N/V	
VS IIITM	M	N/V	N/V	N/V	N/	N/V	N/V	✓
Astral™	N/V	N/V	N/V	N/V	V	FULL FACE	FULL FACE	✓
	FUNL FACE				N/			N/V

¹ SmartStop may not operate as described when using this mask with some CPAP or bi-level devices. ² Excluding PAC and iVAPS modes. Only compatible for pressure range 4 - 25 cm H2O.

N/

V

N/

N/

N/

Full Face Masks 2

V	V
---	---

Device	N/ V	Mirage™	Mirage Series 2	Ultra Mirage ^ℕ ⁄ ∨	Mirage Quattro™	Quattro™ FX / Quattro™ FX fo Quattro™ Air / Quattro™ Air fo		Hospital Full Face
AirCurve™ 11 VAuto / AirCurve™ 11 S / AirCurve™ 11	N/ STV	N/V	N/V	Full Face V	Full Face	Full Face Full Face	Full Face	Full Face
AirCurve™ 11 ASV PaceWave	N/	N/V	N/V	Full Face N/	Full Face	Full Face	Full Face	Full Face
AirSense™ 11	V	N/V	N/V	Full Face V	Full Face	Y	Full Face	Full Face
AirMini™		Χ	Х	X N/ V	X	^	X	X

³ This combination of mask and device has been determined safe for use based on the evaluation of accuracy of monitoring of pressure, volume, minute ventilation and respiratory rate, and its alarms if applicable. Small variations in performance might be observed, As per expected clinical practice, the effectiveness of ventilation and alarms should be verified by the clinician.

⁴SmartStart may not operate as described when using this mask with some CPAP or bi-level devices.

N/V = The mask has not been validated for this device.

 $[\]checkmark$ = Compatible X = Not Compatible

Device	Mirage™	Mirage Series 2	Ultra Mirage™	Mirage Quattro™	Quattro™ FX / Quattro™ FX fo Quattro™ Air / Quattro™ Air fo	r Her Mirage Liberty™ r Her	Hospital Full Face
AirStart™ 10 / AirSense™ 10	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve TM 10 VAuto / AirCurve TM 10 S /	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 ST	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 CS Pacewave	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AirCurve™ 10 CS-A Pacewave	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
Lumis™	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
S9 AutoSet™ / S9 Elite™	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
S9 Escape™	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE1	FULL FACE
AutoSet CS™ 2	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AutoSet CS™ (S9)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AutoSet CS™ (S9 PaceWave)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
AutoSet CS-A™ (S9				Jo setting required Consu	ult your clinician before changing to a c	different mask type	
Page Mayo)					ult your clinician before changing to a	· · · · · · · · · · · · · · · · · · ·	
PaceWave) C-Series Tango™	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR	MIR
S7™ Lightweight	STANDARD	ULTRA	MIRAGE	MIRAGE	MIRAGE	FULL	FULL
S7 Elite TM / S7 AutoSet Spirit	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR	MIR
•	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	FULL	FULL
S7 AutoSet Respond™ S8™ Auto 25	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR	MIR
S8 AutoSet Spirit TM			Pog		at 12 cm H2O. Consult your clinician a	and shook the transquide	FULL
S8 AutoSet Spint *** S8 AutoSet Vantage TM					at 12 cm H2O. Consult your clinician a		MIR
S8 Escape TM	MIR	MIR	MIR	MIR FULL	MIR FULL	MR ^L	FHAL
S8 Escape™ II	FULL	FULL	FULL	MIR FULL	MIR FULL	FUEL	FUEL
S8 Elite TM / S8 Elite TM II	MIR	MIR			ult your clinician before changing to a		FMAL
S8 AutoSet Spirit™ II	N	NNNNNNNNNNNNNNNNNNNNNNNNNNNNNNNN	NEMARIFULL	MIR FULL	MIR FULL	MH-FULL	MHRFULL
S8 Compact TM	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
VPAPTM III / VPAPTM III ST	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
VPAP™ III ST-A	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE1	FULL FACE
VPAP™ IV / VPAP™ IV ST	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAP™ Adapt SV	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAPTM Adapt (S9)	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE
VPAP™ Adapt (S9 PaceWave)	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
-VPAP TM Malibu	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
-VPAP™ Auto 25	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL	MIR FULL
-Auto 25 (S9)	N/V	N/V	Full Face			Full Face	Full Face
VPAPTM S / VPAPTM ST	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAPTM S (S9) / VPAPTM ST (S9)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
(S9) VPAP™ ST-A (S9)	N/V	N/V	Full Face	Full Face	Full Face	Full Face	Full Face
VPAPTM ST (S9 iVAPS) /	N/V	N/V	FULL FACE	Full Face	Full Face	FULL FACE	FULL FACE
VPAPTM ST-A (S9 iVAPS) S9	N/V	N/V	FULL FACE	FULL FACE	FULL FACE	FULL FACE	FULL FACE
VPAP™ Tx VPAP™ Tx Stellar™				FULL FACE	FULL FACE		

Device	Mira	ge™	Mirage Ser	ies 2	Ultra Mirage™	Mirage Quattro™	M Quattr Quattr	ro™ FX / Quattro ro™ Air / Quattro	™ FX for Her N ™ Air for Her	Mirage Liberty [™]	M Hosp	oital Full Face
VS Ultra VS Integra™	N/		N/		✓	✓		N/V		~		<u> </u>
VS IĬŤ	V		V		✓	✓		N/V		✓		✓
Astra™	N/		N/		FULL FACE	FULL FACE		FULL FACE		FULL FACE	F	ULL FACE
	V		V									
NI I NA I . A	N/		N/									
Nasal Masks 1	V		V									
Device	AirTouch [™] N30i ¹	[™] AirFit [™] Ni for AirMin	30AirFit™ N30i i AirFit™ N30	1A⁄irFit™ N) Classic	J20 AirFit™ N20 / AirFit™ N20 for I	AirTouch™ N20 / H Ai rTouch™ N20 for	/ AirFit™ N H Ai rFit™ N for Her	10/Swift™ FX N 10 Swift™ FX N for Her	andMirage™ FX andMirage FX™ Her	⟨ / Mirage foActiva™ LT	Mirage™ SoftGel	Hospital Nasal
AirCurve™ 11 VAuto / AirCurve™ 11 S / AirCurve™	Pillows	Х	Pillows	Pillo	Pillo	Pillow	Pillows Pillows	Pillows	Nasa	Nasa	Χ	Nasal
11 ST	Pillows	Χ	Pillows	WS	WS	S	Pillows	Pillows	I	I	Χ	Nasal
AirCurve™ 11 ASV PaceWave	Pillows	Χ	Pillows	Pillo	Pillo	Pillow	F IIIOVVS	Pillows	Naca	Nasa	N/V	Nasal
AirSense™ 11	Χ	✓	Χ	→ FIII0 ✓	Y	- FIIIOW	Χ	X	———Nasa—— X	X	Χ	X
AirMini TM	Pillows	Χ	Pillows	Millows	Millows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AirStart [™] 10 / AirSense [™] 10 AirCurve [™] 10 VAuto /	Pillows	Χ	Pillows	Pillo Pillows	Pillo Pillo	Pillows	Pillows	Pillows	Nasa Nasal	Nasa Nasal	Nasal	Nasal
AirCurve™ 10 S /	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AirCurve™ 10 ST	N/V	Χ	N/V	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AirCurve™ 10 CS Pacewave	Pillows ³	Χ	Pillows3	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AirCurve™ 10 CS-A Pacewave	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
Lumis™	Pillows	X	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal ——
S9 AutoSet™ / S9 Elite™	N/V	X	N/V	N/V	N/V	N/V	N/V	N/V	VISTA1	ACTIVA	ACTIVA	VISTA
S9 Escape™	N/V	X	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AutoSet CS™ 2	N/V	X	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AutoSet CS™ (S9)	N/V	X	N/V	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
AutoSet CS™ (S9 PaceWave)	N/V	X	N/V	N/V	N/V	N/V	N/V					
AutoSet CS-A™ (S9	N/V	X	N/V	N/V	N/V	N/V	N/V	N/V	No setting req	uired. Consult you	ır clinician bef	ore changing to a
D===\\\/\	N/V	X	N/V	N/V	N/V	N/V	N/V	N/V	different mask	type. No setting	required. Cor	sult your clinician
PaceWave)	N/V	X	N/V	N/V	N/V	N/V	N/V	N/V	befgteathenaire	D to a STIFACHEDIA RIBS	k s yttændard	STANDARD
C-Series Tango™	N/V	Χ	N/V	N/V	SWIFT	SWIFT	SWIFT	N/V	STANDARD	STANDARD	STANDARD	STANDARD
S7TM Lightweight	N/V	Χ	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD) ACTIVA	ACTIVA	STANDARD
S7 Elite™ / S7 AutoSet Spirit	N/V	X	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD) ACTIVA	ACTIVA	STANDARD
S7 AutoSet Respond™	N/V	X	N/V	N/V	N/V			SWIFT	STANDARD) ACTIVA	ACTIVA	STANDARD
S8™ Auto 25 S8 AutoSet Spirit™	N/V	X	N/V	N/V	.,,,	N/V	Red	quires pressure ca	libration at 12 cm	H2O. Consult yo	our clinician a	nd check the user g
•	N/V	Χ	N/V	N/V		Requires pre		ation at 12 cm H20				
S8 AutoSet Vantage TM	N/V	X	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARE		ACTIVA	STANDARD
S8 Escape™ S8 Escape™ II	N/V	X	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARE		ACTIVA	STANDARD
S8 Elite TM / S8 Elite TM II	N/V	X	N/V	N/V	N/V	N/V						ifferent mask type.
	N/V	X	N/V	N/V	N/V	N/V	MIRAGE		STANDARI			
S8 AutoSet Spirit TM II	N/V	X	N/V	N/V	N/V	N/V	MIRAGE		STANDARI			
S8 Compact TM	N/V	X	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARI		ACTIVA	STANDARD
VPAP™ III / VPAP™ III ST VPAP™ III ST-A					N/V	N/V	N/V	N/V	VISTA1	ACTIVA	ACTIVA	VISTA
					•	1 4/ V	1 N/ V	1 V/ V	VIOIAI	AOTIVA	,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,	
VPAPTM IV / VPAPTM IV ST												English 4

VPAP™ Adapt SV

Device	AirTouch ¹ N30i ¹	^M AirFit™ N3 for AirMini	ØirFit™ N30 AirFit™ N3	i1A⁄irFit™ N2 0 Classic		AirTouch™ N20 l&i rTouch™ N20 fo				Mirage pActiva™ LT	Mirage™ SoftGel	Hospital Nasal
VPAP™ Adapt (S9)	N/V	X	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAP™ Adapt (S9	N/V	X	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
PaceWave) VPAP TM	N/V	X	N/V	N/V	N/V	N/V	SWIFT	SWIFT	VISTA1	N/V	N/V	MERIDIAN
Malibu VPAP™ Auto 25	N/V	Χ	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
Auto 25 (S9) VPAPTM S /	N/V	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAPTM ST VPAPTM S	N/V	X	N/V	N/V	SWIFT	SWIFT	SWIFT	SWIFT	STANDARD	ACTIVA	ACTIVA	STANDARD
(S9) / VPAPTM ST (S9)	Pillows	X	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAPTM ST-A (S9)	N/V	X	N/V	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAP™ ST (S9 iVAPS) / VPAP™ ST-A (S9	N/V	X	N/V	Pillows	Pillows	Pillows	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
VPAPTM ST-A (S9 iVAPS)	Pillows2	X	Pillows2	N/V	Pillows	N/V	Pillows	Pillows	Nasal	Nasal	Nasal	Nasal
S9 VPAP™ Tx	N/V	×	N/V	N/V	NA	N/V	PILLOWS	PILLOWS	NASAL	NASAL	NASAL	NASAL NASAL
VPAP™ Tx	N/V	×	N/V	PILLOWS	PILLOWS	PILLOWS	PILLOWS1	PILLOWS1	NASAL	NASAL	NASAL	NASAL
Stellar™	N/V	Y	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	NV
VS Ultra™ / VS Integra™	N/V	X	N/V	N/V	N/V	N/V	N/V	N/V	N/V	1 N/ V	N/V	1 N/ V
_VS III™	N/V	X	N/V	Pillows	Pillows	Pillows	N/V	N/V	Nasal	~	Nasal	~
Astral™		.,	.,,							N/V		Nasal

¹ SmartStop may not operate as described when using this mask with some CPAP or bi-level devices.

Nasal Masks 2

Device	Pixî™	Mirage Kidsta™	Mirage Micro™	Mirage Vista™	Ultra Mirage II	Mirage	Mirage Activa™	Modular	Papillon™	Meridian
AirCurve™ 11 VAuto /		X	X	X	X	N/V	X	N/V	N/V	X
AirCurve™ 11 S / AirCurve™ 11 ST	Pediatric				,,	, .		, .	, .	
_AirCurve™ 11 ASV PaceWave		X	X	X	X	N/V	X	N/V	N/V	X
AirSense™ 11	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V
AirMini™	N/V	X	X	X	X	X	X	X	X	X
AirStart™ 10 / AirSense™ 10	X	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
AirCurve TM 10 VAuto / AirCurve TM 10	s/N/V	Nasal	Nasal	Nasal	Nasal	NA	Nasal	NV	N/V	Nasal
AirCurve™ 10 ST	N/V					•		,		
AirCurve™ 10 CS Pacewave	NIA/	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
AirCurve™ 10 CS-A Pacewave	N/V	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
Lumis™	N/V	Nasal	Nasal	Nasal	Nasal	N/V	Nasal	N/V	N/V	Nasal
S9 AutoSet™ / S9 Elite™	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
S9 Escape™	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
AutoSet CS™ 2	N/V	N/V	VISTA1	VISTA	ULTRA	N/V	ACTIVA	N/V	PAPILLON	VISTA
AutoSet CS™ (S9)	N/V N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal

 $^{^{2}}$ Excluding PAC and iVAPS modes. Only compatible for pressure range 4 - 25 cm H2O.

³ This combination of mask and device has been determined safe for use based on the evaluation of accuracy of monitoring of pressure, volume, minute ventilation and respiratory rate, and its alarms if applicable. Small variations in performance might be observed, As per expected clinical practice, the effectiveness of ventilation and alarms should be verified by the clinician.

N/V =The mask has not been validated for this device.

 $[\]checkmark$ = Compatible X = Not Compatible

Device AutoSet CS™ (S9	Pixî™	Mirage Kidsta™	Mirage Micro™	•	Ultra Mirage II	Mirage	Mirage Activa™	Modular	Papillon™	Meridian
PaceWave) / AutoSet CS- A TM (S9 PaceWave)	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
C-Series Tango™	N/V	N/V	N/V		No settina require	ed. Consult vou	clinician before chang	ing to a different r	mask type.	N/V N/V
S7™ Lightweight	N/V	N/V	N/V			•	clinician before chang			N/V
S7 Elite™ / S7 AutoSet	N/V	N/V N/V	STANDARD STANDARD	STANDA	ULTRA	MIRAGE	STANDARD	STANDARD	STANDARD	N/V
Spirit	N/V	N/V	STANDARD	RD	ULTRA	MIRAGE	STANDARD	STANDARD	STANDARD	STANDARD
S7 AutoSet Respond™	N/V	N/V	STANDARD	STANDA	ULTRA	MIRAGE	ACTIVA	N/V	N/V	STANDARD
S8 TM Auto 25	N/V	N/V	STANDARD	RD	ULTRA	MIRAGE	ACTIVA	STANDARD	STANDARD	STANDARD
S8 AutoSet Spirit™	N/V	N/V	STANDAND	STANDA	ULTRA	MIRAGE	ACTIVA	STANDARD	STANDARD	STANDAND
S8 AutoSet Vantage™	N/V	N/V		RD	Requires pressure ca	alibration at 12 c	m H2O. Consult your o	clinician and check	the user guide.	
S8 Escape™	N/V	N/V		STANDA			m H2O. Consult your o			
S8 Escape™ II	N/V	N/V	STANDA	8 P anda	ULTRA	MIRAGE	ACTIVA	STANDARD	STANDA	STANDA
S8 Elite™ / S8 Elite™ II	N/V	N/V	RD	STANDA	ULTRA	MIRAGE	ACTIVA	STANDARD	RD	RD
S8 AutoSet Spirit™ II	N/V	N/V	STANDA	8 P anda	No setting require	ed. Consult your	clinician before chang	ing to a different r	masiSTANDA	STANDA
S8 Compact™	N/V	STANDARD	SP ANDARD	SPANDARD	ULTRA	MIRAGE	STANDARD	STANDARD	RDN/V	RD _{N/V}
VPAPTM III / VPAPTM III ST	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	STANDARD	STANDARD	N/V	N/V
VPAP™ III ST-A	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	N/V	N/V	STANDARD
VPAPTM IV / VPAPTM IV ST	N/V	N/V	VISTA1	VISTA	ULTRA	N/V	ACTIVA	N/V	N/V	VISTA
-VPAP™ Adapt SV	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
VPAP™ Adapt (S9)	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
-VPAP TM Adapt (S9 PaceWave)	N/V	N/V	MERIDIAN	VISTA	ULTRA	N/V	N/V	N/V	N/V	MERIDIAN
VPAP™ Malibu	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	N/V	N/V	STANDARD
VPAP™ Auto 25	N/V	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	STANDARD
Auto 25 (S9)	N/V	N/V	STANDARD	STANDARD	ULTRA	MIRAGE	ACTIVA	N/V	N/V	STANDARD
VPAPTM S / VPAPTM ST	Pediatric	<u>.</u>	Nasal	Nasal	N = = = 1 1 11±==	NIA /	NI	Nasal	Nasal	Nasal
VPAPTM S (S9) / VPAPTM					Nasal Ultra	N/V	Nasal			
ST (S9) / - VPAP TM ST-A (S9)	Pediatric	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	Nasal	Nasal	Nasal
VPAP™ ST (S9 iVAPS) /	Pediatric	N/V	Nasal	Nasal	Nasal Ultra	N/V	Nasal	N/V	N/V	Nasal
VPAP™ ST-A (S9 iVAPS)	N/V	NASAL	NASAL	NASAL	NASAL-ULTRA	N/V	NASAL	N/V	N/V	NASAL
S9 VPAP™ Tx	Pediatric	NASAL	NASAL	NASAL	ULTRA	N/V	NASAL	N/V	N/V	N/V
VPAP™ Tx	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V	N/V
Stellar™	N/V	✓	✓	~	✓	N/V	✓	N/V	N/V	N/V
VS Ultra™ / VS Integra™	Pediatric	Nasal	Nasal	Nasal	Nasal Ultra	N/V	N/V	N/V	N/V	N/V

VS IIITM
N/V = The mask has not been validated for this device.
Astral IM
= Compatible X = Not Compatible

Nasal Pillows

Device	AirFit™30i1	AirFit™ P10 For AirMini™	AirFit™ P10 / AirFit™ P10 for Hei	Mirage Swift™ r	Mirage Swift∏ ^M	Swift™ FX / Swift™ FX for Her / Swift™ FX Bella / Swift™ FX Bella (Swift™T Gray	Swift™ LT for Her/ Swift™ LT-F
AirCurve™ 11 VAuto / AirCurve™ 11 S / AirCurve™	Pillows	X	Pillows2	N/V	N/V	Pillows	Pillows	Pillows
11 ST	Pillows	X	Pillows2	N/V	N/V	Pillows	Pillows	Pillows
AirCurve [™] 11 ASV PaceWave	Pillows	X	Pillows	N/V	Pillows	Pillows	Pillows	Pillows
AirSense™ 11		,	X	<u> </u>		V		X
AirMini™	X	<u> </u>	Pillows2	X	X	X	X	Pillows
AirStart™ 10 / AirSense™ 10	Pillows	X	Pillows2	N/V	Pillows	Pillows	Pillows	Pillows
AirCurve™ 10 VAuto /	Pillows	X	Pillows2	N/V	Pillows	Pillows	Pillows	Pillows
AirCurve™ 10 S /	Pillows	Χ	Pillows2	N/V	Pillows	Pillows	Pillows	Pillows
AirCurve™ 10 ST	N/V	Χ	Pillows2	N/V	Pillows	Pillows	Pillows	Pillows
AirCurve™ 10 CS Pacewave	Pillows4	Χ	Pillows	N/V	Pillows	Pillows	Pillows	Pillows
AirCurve™ 10 CS-A Pacewave	Pillows	Χ	Pillows	Pillows	Pillows	Pillows	Pillows	Pillows
-Lumis TM	Pillows	X	N/V	Pillows	Pillows	Pillows	Pillows	N/V
S9 AutoSet™ / S9 Elite™	N/V	X	Pillows2	N/V	N/V	N/V	N/V	Pillows
S9 Escape™	Pillows	X	Pillows2	Pillows	Pillows	Pillows	Pillows	Pillows
AutoSet CS™ 2	Pillows	X	Pillows2	Pillows	Pillows	Pillows	Pillows	
AutoSet CS™ (S9)	N/V	X	1 11104432	Pillows	Pillows	Pillows	Pillows	Pillows
AutoSet CS™ (S9 PaceWave)	N/V	X						
AutoSet CS-A™ (S9	N/V	X			No setting required. C	onsult your clinician before changing to a d	lifferent mask type.	
	N/V	X			No setting required. C	onsult your clinician before changing to a d	lifferent mask type.	
PaceWave)	N/V	X	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE
C-Series Tango™	N/V	X	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE
S7™ Lightweight	N/V	X	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
S7 Elite™ / S7 AutoSet Spirit	N/V	X	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
S7 AutoSet Respond™	N/V	X	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
S8™ Auto 25	N/V	X		Rec	uires pressure calibra	tion at 12 cm H2O. Consult your clinician a		iide
S8 AutoSet Spirit™	N/V	X				tion at 12 cm H2O. Consult your clinician a		
S8 AutoSet Vantage™	N/V		SWI	SWI	SWIFT	SWIFT	SWIFT	SWI
S8 Escape™	N/V	X	FT	FT				FT
S8 Escape™ II	N/V	X	SWI		SWIFT	SWIFT	SWIFT	SWI
S8 Elite™ / S8 Elite™ II	N/V	X				onsult your clinician before changing to a d		
S8 AutoSet Spirit™ II	N/V	X	MīRAGE 2	MTRAGE	MIRAGE	MIRAGE	MIRAGE	MTRAGE
S8 Compact™	N/V	X	MIRAGE 2	MIRAGE	MIRAGE	MIRAGE	MIRAGE	MIRAGE
VPAPTM III / VPAPTM III ST	Pillows	X	SWIFT2	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
VPAP™ III ST-A	Pillows	X	N/V	N/V	N/V	N/V	N/V	N/V
VPAPTM IV / VPAPTM IV ST	N/V	Χ	Pillows2	Pillows	Pillows	Pillows	Pillows	Pillows
VPAP™ Adapt SV	N/V	Χ	Pillows2	Pillows	Pillows	Pillows	Pillows	Pillows
VPAP™ Adapt (S9)	Pillows	Χ	SWIFT2	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
VPAP™ Adapt (S9 PaceWave)	N/V	Χ	SWIFT2	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
VPAP TM Malibu	Pillows	Χ	Pillows2	Pillows	Pillows	Pillows	Pillows	Pillows
VPAP™ Auto 25	1 11104/2	X	SWIFT2	SWIFT	SWIFT	SWIFT	SWIFT	SWIFT
Auto 25 (S9)			Pillows2	Pillows	Pillows	Pillows	Pillows	Pillows
Auto 25 (39)								

VPAPTM S / VPAPTM ST

VPAPTM S (S9) / VPAPTM ST

(S9)

Device	AirFit™30i1	AirFit™ P10 For AirMini™	AirFit™ P10 / AirFit™ P10 for He	Mirage Swift™ er	Mirage SwiftI™	Swift™ FX / Swift™ FX for Her / Swift™ FX Bella / Swift™ FX Bella	Swift™T Gray	Swift™ LT for Her/ Swift™ LT-F
VPAP™ ST-A (S9)	N/V	X	Pillowŝ	Pillows	Pillows	Pillows	Pillows	Pillows
VPAP™ ST (S9 iVAPS) /	N/V	Х	Pillows2	Pillows	Pillows	Pillows	Pillows	Pillows
VPAPTM ST-A (S9 iVAPS)	Pillows3	Χ	Pillows2	N/V	Pillows	Pillows	Pillows	Pillows
S9 VPAP™ Tx	N/V	X	PILLOWS2	PILLOWS	PILLOWS	PILLOWS	PILLOWS	PILLOWS
VPAP™ Tx	N/V	Χ	PILLOWS1, 2	PILLOWS1	PILLOWS1	PILLOWS1	PILLOWS1	PILLOWS1
Stellar™	N/V	X	N/V	N/V	N/V	N/V	N/V	N/V
	N/V	X	N/V	N/V	N/V	N/V	N/V	N/V
VS Ultra™ / VS Integra™	N/V	X	N/V	Pillows	Pillows	Pillows	Pillows	Pillows

VS III

N/V = The mask has not been validated for this device.

 \checkmark = Compatible X = Not Compatible

¹SmartStop may not operate as described when using this mask with some CPAP or bi-level devices.

²Accuracy of tidal volume/minute ventilation may reduce using the AirFit P10 mask in bilevel modes, particularly at higher pressure supports.

³Excluding PAC and iVAPS modes. Only compatible for pressure range 4 - 25 cm H2O.

⁴ This combination of mask and device has been determined safe for use based on the evaluation of accuracy of monitoring of pressure, volume, minute ventilation and respiratory rate, and its alarms if applicable. Small variations in performance might be observed, As per expected clinical practice, the effectiveness of ventilation and alarms should be verified by the clinician.