



**ResMed**

**AirSense™ 10** AUTOSET

AUTOSET FOR HER

ELITE

CPAP

 **Respshop**  
better sleep, better life.  
**Click Here To Buy**



User guide  
English

## Welcome

The AirSense 10™ AutoSet™ and AirSense 10 AutoSet for Her are ResMed's premium auto-adjusting pressure devices. The AirSense 10 Elite and AirSense 10 CPAP are ResMed's Continuous Positive Airway Pressure (CPAP) devices.

### WARNING

Read this entire guide before using the device.

### CAUTION

In the US, Federal law restricts this device to sale by or on the order of a physician.

## Indications for use

### AirSense 10 AutoSet

The AirSense 10 AutoSet self-adjusting device is indicated for the treatment of obstructive sleep apnea (OSA) in patients weighing more than 66 lb (30 kg). It is intended for home and hospital use.

The humidifier is intended for single patient use in the home environment and re-use in a hospital/institutional environment.

### AirSense 10 AutoSet for Her

The AirSense 10 AutoSet for Her self-adjusting device is indicated for the treatment of obstructive sleep apnea (OSA) in patients (female patients with mild to moderate OSA when using AutoSet for Her treatment mode) weighing more than 66 lb (30 kg). It is intended for home and hospital use.

The humidifier is intended for single patient use in the home environment and re-use in a hospital/institutional environment.

### AirSense 10 Elite

The AirSense 10 Elite device is indicated for the treatment of obstructive sleep apnea (OSA) in patients weighing more than 66 lb (30 kg). It is intended for home and hospital use.

The humidifier is intended for single patient use in the home environment and re-use in a hospital/institutional environment.

### AirSense 10 CPAP

The AirSense 10 CPAP device is indicated for the treatment of obstructive sleep apnea (OSA) in patients weighing more than 66 lb (30 kg). It is intended for home and hospital use.

The humidifier is intended for single patient use in the home environment and re-use in a hospital/institutional environment.

## Contraindications

Positive airway pressure therapy may be contraindicated in some patients with the following pre-existing conditions:

- severe bullous lung disease
- pneumothorax
- pathologically low blood pressure
- dehydration
- cerebrospinal fluid leak, recent cranial surgery, or trauma.

## Adverse effects

You should report unusual chest pain, severe headache, or increased breathlessness to your prescribing physician. An acute upper respiratory tract infection may require temporary discontinuation of treatment.

The following side effects may arise during the course of therapy with the device:

- drying of the nose, mouth, or throat
- nosebleed
- bloating
- ear or sinus discomfort
- eye irritation
- skin rashes.

## At a glance


The AirSense 10 includes the following:

- Device with HumidAir™ integrated humidifier
- Water tub
- Air tubing
- Power supply unit
- Travel bag
- SD card (not available in all devices).

Contact your care provider for a range of accessories available for use with the device including:

- Air tubing (heated and non-heated): ClimateLineAir™, SlimLine™, ClimateLineAir Oxy, Standard
- Water tub: Standard water tub, cleanable water tub (can be disinfected)
- Side cover for use without the humidifier
- Filter: Hypoallergenic filter, standard filter
- Air10™ DC/DC converter
- SD card reader
- Air10 oximeter adapter
- Air10 USB adapter
- Power Station II
- Air10 tubing elbow.


About your device




- 1 Air outlet
- 2 Air filter cover
- 3 Power inlet
- 4 Serial number and device number

- 5 Water tub
- 6 Screen
- 7 Adapter cover
- 8 SD card cover

About the control panel

 Start/Stop button

 Dial










 Home button

Press to start/stop therapy.  
Press and hold for three seconds to enter power save mode.







Turn to navigate the menu and press to select an option.  
Turn to adjust a selected option and press to save your change.

Press to return to the Home screen.

Different icons may be displayed on the screen at different times including:

- |   |                    |   |                                      |
|---|--------------------|---|--------------------------------------|
|  | Ramp Time          |  | Wireless signal strength (green)     |
|   | Ramp Time Auto     |  | Wireless transfer not enabled (gray) |
|  | Humidity           |  | No wireless connection               |
|  | Humidifier warming |  | Airplane Mode                        |
|  | Humidifier cooling |   |                                      |

## Setup



### CAUTION

**Do not overfill the water tub as water may enter the device and air tubing.**

1. Place the device on a stable level surface.
2. Plug the power connector into the rear of the device. Connect one end of the power cord into the power supply unit and the other end into the power outlet.
3. Connect the air tubing firmly to the air outlet located on the rear of the device.
4. Open the water tub and fill it with distilled water up to the maximum water level mark. Do not fill the water tub with hot water.
5. Close the water tub and insert it into the side of the device.
6. Connect the free end of the air tubing firmly onto the assembled mask. See the mask user guide for detailed information.

Recommended masks are available on [www.resmed.com](http://www.resmed.com).

## Starting therapy

1. Fit your mask.
2. Press Start/Stop or breathe normally if SmartStart™ is enabled.

You will know that therapy is on when the **Sleep Report** screen is displayed.



The current treatment pressure is shown in green.

During ramp time the pressure is gradually increasing and you will see a spinning circle. Once the prescribed treatment pressure is reached, the entire circle will be green.

The screen will go black automatically after a short period of time. You can press Home or the dial to turn it back on. If power is interrupted during therapy, the device will automatically restart therapy when power is restored.

The AirSense 10 device has a light sensor that adjusts the screen brightness based on the light in the room.

## Stopping therapy

1. Remove your mask.
2. Press Start/Stop or if SmartStart is enabled, therapy will stop automatically after a few seconds.

The **Sleep Report** now gives you a summary of your therapy session.



**Usage hours**—Indicates the number of hours of therapy you received last session.

**Mask Seal**—Indicates how well your mask sealed:

😊 Good mask seal.

😞 Needs adjusting, see Mask Fit.

**Humidifier**—Indicates if your humidifier is working properly:

😊 Humidifier working.

😞 Humidifier might be faulty, contact your care provider.

If set by your care provider, you will also see:

**Events per hour**—Indicates the number of apneas and hypopneas experienced per hour.

**More Info**—Turn the dial to scroll down to view more detailed usage data.

## Power save mode

Your AirSense 10 device records your therapy data. In order to allow it to transmit the data to your care provider, you should not unplug the device. However, you can put it into power save mode to save electricity.

To enter power save mode:


- Press and hold Start/Stop for three seconds.  
The screen goes black.

To exit power save mode:

- Press Start/Stop once.  
The **Home** screen is displayed.

## My Options

Your AirSense 10 device has been set up for your needs by your care provider, but you may find you want to make small adjustments to make your therapy more comfortable.




Highlight **My Options** and press the dial to see your current settings. From here, you can personalize your options.

## Ramp Time

Designed to make the beginning of therapy more comfortable, Ramp Time is the period during which the pressure increases from a low start pressure to the prescribed treatment pressure.

You can set your Ramp Time to Off, 5 to 45 minutes or Auto. When Ramp Time is set to **Auto**, the device will detect when you have fallen asleep and then automatically rise to the prescribed treatment pressure.



To adjust Ramp Time:

1. In **My Options**, turn the dial to highlight **Ramp Time** and then press the dial.
2. Turn the dial to adjust the ramp time to your preferred setting and press the dial to save the change.

## Humidity Level

The humidifier moistens the air and is designed to make therapy more comfortable. If you are getting a dry nose or mouth, turn up the humidity. If you are getting any moisture in your mask, turn down the humidity.

You can set the Humidity Level to Off or between 1 and 8, where 1 is the lowest humidity setting and 8 is the highest humidity setting.



### To adjust the Humidity Level:

1. In **My Options**, turn the dial to highlight **Humidity Level** and then press the dial.
2. Turn the dial to adjust the humidity level and press the dial to save the change.

If you continue to get a dry nose or mouth, or moisture in your mask, consider using ClimateLineAir heated air tubing. ClimateLineAir together with Climate Control delivers more comfortable therapy.

## Mask Fit

Mask Fit is designed to help you assess and identify possible air leaks around your mask.



### To check Mask Fit:

1. Fit the mask as described in the mask user guide.
2. In **My Options**, turn the dial to highlight **Run Mask Fit** and then press the dial.  
The device starts blowing air.
3. Adjust the mask, mask cushion and headgear until you get a **Good** result.

To stop Mask Fit, press the dial or Start/Stop. If you are unable to get a good mask seal, talk to your care provider.

## More options

There are some more options on your device which you can personalize.

**Pressure Relief\*** When Pressure Relief is enabled, you may find it easier to breathe out. This can help you get used to therapy.

**SmartStart\*** When SmartStart is enabled, therapy starts automatically when you breathe into your mask. When you remove your mask, it stops automatically after a few seconds.

\*When enabled by your care provider.

## Caring for your device





It is important that you regularly clean your AirSense 10 device to make sure you receive optimal therapy. The following sections will help you with disassembling, cleaning, checking and reassembling your device.



## WARNING

Regularly clean your tubing assembly, water tub and mask to receive optimal therapy and to prevent the growth of germs that can adversely affect your health.

### Disassembling



1. Hold the water tub at the top and bottom, press it gently and pull it away from the device.
2. Open the water tub and discard any remaining water.
3. Hold the cuff of the air tubing and gently pull it away from the device.
4. Hold both the cuff of the air tubing and the swivel of the mask, then gently pull apart.

### Cleaning

You should clean the device weekly as described. Refer to the mask user guide for detailed instructions on cleaning your mask.



1. Wash the water tub and air tubing in warm water using mild detergent.  
Do not wash in a dishwasher or washing machine.
2. Rinse the water tub and air tubing thoroughly and allow to dry out of direct sunlight and/or heat.
3. Wipe the exterior of the device with a dry cloth.

## Checking

You should regularly check the water tub, air tubing and the air filter for any damage.

1. Check the water tub:
  - Replace it if it is leaking or has become cracked, cloudy or pitted.
  - Replace it if the seal is cracked or torn.
  - Remove any white powder deposits using a solution of one part household vinegar to 10 parts water.
2. Check the air tubing and replace it if there are any holes, tears or cracks.
3. Check the air filter and replace it at least every six months. Replace it more often if there are any holes or blockages by dirt or dust.

To replace the air filter:



1. Open the air filter cover and remove the old air filter.  
The air filter is not washable or reusable.
2. Place a new air filter onto the air filter cover and then close it.  
Make sure the air filter is fitted at all times to prevent water and dust from entering the device.

## Reconnecting the air tubing


When the water tub and air tubing are dry, you can reassemble the parts.

1. Connect the air tubing firmly to the air outlet located on the rear of the device.
2. Open the water tub and fill it with distilled room temperature water up to the maximum water level mark.
3. Close the water tub and insert it into the side of the device.
4. Connect the free end of the air tubing firmly onto the assembled mask.

## Therapy data

Your AirSense 10 device records your therapy data for you and your care provider so they can view and make changes to your therapy if required. The data is recorded and then transferred to your care provider wirelessly, if wireless network is available, or via an SD card.

## Wireless

Your AirSense 10 device is equipped with cellular communication. This allows your therapy data to be wirelessly transmitted to your care provider to enhance the quality of your treatment. It also allows therapy settings to be updated in a more timely manner or your device software to be upgraded. The Wireless signal strength icon  displayed at the top right of your screen indicates the signal strength.

The data is usually transmitted after therapy has stopped. In order to make sure that your data is transferred, leave your device connected to the power outlet at all times and make sure that it is not in Airplane Mode.

#### Notes:

- Therapy data might not be transmitted if you use it outside of the country or region of purchase.
- Devices with cellular communication might not be available in all regions.


Please be aware that within the wireless network, the availability and quality of the network may be affected by terrain, buildings and the weather. Wireless communication depends on network availability. Coverage is not available everywhere and varies by service.

## SD card

An alternative way for your therapy data to be transferred to your care provider is via the SD card, if provided. Your care provider may ask you to send the SD card by mail or to bring it in. When instructed by your care provider, remove the SD card.

Do not remove the SD card from the device when the SD light is flashing, because data is being written to the card.

#### To remove the SD card:



1. Open the SD card cover.
2. Push in the SD card to release it. Remove the SD card from the device.  
Place the SD card in the protective folder and send it back to your care provider.

For more information on the SD card refer to the SD card protective folder provided with your device.

**Note:** The SD card should not be used for any other purpose.

## Traveling

You can take your AirSense 10 device with you wherever you go. Just keep the following in mind:

- Use the travel bag provided to prevent damage to the device.
- Empty the water tub and pack it separately in the travel bag.
- Make sure you have the appropriate power cord for the region you are traveling to. For information on purchasing, contact your care provider.
- If you are using an external battery, you should turn off the humidifier in order to maximize the life of your battery. Do this by turning the **Humidity Level** to Off.


# Traveling by plane

Your AirSense 10 device may be taken on board as carry-on luggage. Medical devices do not count toward your carry-on luggage limit.

You can use your AirSense 10 device on a plane as it meets the Federal Aviation Administration (FAA) requirements. Air travel compliance letters can be downloaded and printed from [www.resmed.com](http://www.resmed.com).

When using the device on a plane:

- Make sure the water tub is completely empty and inserted into your device. The device will not work without the water tub inserted.
- Turn on **Airplane Mode**.



## To turn on Airplane Mode:

1. In **My Options**, turn the dial to highlight **Airplane Mode** and then press the dial.
2. Turn the dial to select **On** and then press the dial to save the change.

The Airplane Mode icon ✈️ is displayed at the top right of the screen.

## ⚠️ CAUTION



Do not use the device with water in the water tub on a plane due to the risk of inhalation of water during turbulence.

# Troubleshooting



If you have any problems, have a look at the following troubleshooting topics. If you are not able to fix the problem, contact your care provider or ResMed. Do not try to open the device.

## General troubleshooting

| Problem/possible cause  | Solution   |
|---|--|
| <b>Air is leaking from around my mask</b><br><br>Mask may be fitted incorrectly.  | Make sure your mask is fitted correctly. See your mask user guide for fitting instructions or use the Mask Fit function to check your mask fit and seal. |
| <b>I am getting a dry or blocked nose</b><br><br>Humidity level may be set too low.                                     | Adjust the Humidity Level.<br><br>If you have ClimateLineAir heated air tubing, see the ClimateLineAir user guide.                                       |
| <b>I am getting droplets of water on my nose, in the mask and air tubing</b><br><br>Humidity level may be set too high. | Adjust the Humidity Level.<br><br>If you have ClimateLineAir heated air tubing, see the ClimateLineAir user guide.                                       |

| Problem/possible cause  | Solution  |
|---|---|
| <b>My mouth is very dry and uncomfortable</b>   |   |
| Air may be escaping through your mouth.   | Increase the Humidity Level.<br><br>You may need a chin strap to keep your mouth closed or a full face mask.  |
| <b>Air pressure in my mask seems too high (it feels like I am getting too much air)</b>   |   |
| Ramp may be turned off.   | Use the Ramp Time option.   |
| <b>Air pressure in my mask seems too low (it feels like I am not getting enough air)</b>  |   |
| Ramp may be in progress.  | Wait for air pressure to build up or turn Ramp Time off.  |
| <b>My screen is black</b>   |   |
| Backlight on the screen may have turned off. It turns off automatically after a short period of time.   | Press Home or the dial to turn it back on.  |
| Power may not be connected.   | Connect the power supply and make sure the plug is fully inserted.  |
| <b>I have stopped therapy, but the device is still blowing air</b>  |   |
| Device is cooling down.   | Device blows a small amount of air in order to avoid condensation in the air tubing. It will stop automatically after 30 minutes.   |
| <b>My water tub is leaking</b>  |   |
| Water tub may not be assembled correctly.   | Check for damage and reassemble the water tub correctly.  |
| Water tub may be damaged or cracked.  | Contact your care provider for a replacement.   |
| <b>My therapy data has not been sent to my care provider</b>  |   |
| Wireless coverage may be poor.  | Make sure that the device is placed where there is coverage (ie, on your bedside table, not in a drawer or on the floor). The Wireless signal strength icon  indicates good coverage when all bars are displayed, and poor coverage when fewer bars are displayed. |
| The No wireless connection icon  is displayed on the top right of the screen. no wireless network available. | Make sure that the device is placed where there is coverage (ie, on your bedside table, not in a drawer or on the floor). If instructed to do so, send the SD card to your care provider. The SD card also contains your therapy data.  |
| Device may be in Airplane Mode.   | Turn off Airplane Mode, see Traveling by plane.   |
| Data transfer is not enabled for your device.   | Talk to your care provider about your settings.   |
| <b>My screen and buttons are flashing</b>   |   |
| Software upgrade is in progress.  | Software upgrade takes approximately 10 minutes to complete.  |



## Device messages

| Device message/possible cause  | Solution   |
|--|--|
| <b>High leak detected, check your water tub, tub seal or side cover</b>                    |  |
| Water tub may not be inserted properly.  | Make sure the water tub is correctly inserted.   |
| Water tub seal may not be inserted properly.   | Open the water tub and make sure that the seal is correctly inserted.  |
| <b>High leak detected, connect your tubing</b>   |  |
| Air tubing may not be connected properly.  | Make sure the air tubing is firmly connected at both ends.   |
| Mask may be fitted incorrectly.  | Make sure your mask is fitted correctly. See your mask user guide for fitting instructions or use the Mask Fit function to check your mask fit and seal.   |
| <b>Tubing blocked, check your tubing</b>   |  |
| Air tubing may be blocked.   | Check the air tubing and remove any blockages. Press the dial to clear the message and then press Start/Stop to restart the device.  |
| <b>SD card error, remove your card and press Start to begin therapy</b>                    |  |
| SD card may not be inserted correctly.   | Remove and reinsert the SD card.   |
| <b>Read only card, please remove, unlock and re-insert SD card</b>                         |  |
| SD card switch may be in the lock (read-only) position.                                    | Move the switch on the SD Card from the lock position  to the unlock position  and then re-insert it. |
| <b>System fault, refer to user guide, Error 004</b>  |  |
| Device may have been left in a hot environment.  | Allow to cool before re-use. Disconnect the power supply and then reconnect it to restart the device.  |
| Air filter may be blocked.   | Check the air filter and replace it if there are any blockages. Disconnect the power supply and then reconnect it to restart the device.   |
| Air tubing may be blocked.   | Check the air tubing and remove any blockages. Press the dial to clear the message and then press Start/Stop to restart the device.  |
| There may be water in the air tubing.  | Empty the water from the air tubing. Disconnect the power supply and then reconnect it to restart the device.  |
| <b>All other error messages, for example, System fault, refer to user guide, Error 0XX</b> |  |
| An unrecoverable error has occurred on the device.   | Contact your care provider. Do not open the device.  |

## Reassembling parts



Some parts of your device are designed to easily come off in order to avoid damage to the parts or the device. You can easily reassemble them as described below.

### To insert the water tub seal:



1. Place the seal into the lid.
2. Press down along all edges of the seal until it is firmly in place.

### To reassemble the water tub lid:



1. Insert one side of the lid into the pivot hole of the base.
2. Slide the other side down the ridge until it clicks into place.

## WARNING

- Make sure that you arrange the air tubing so that it will not twist around the head or neck.
- Make sure the power cord and plug are in good condition and the equipment is not damaged.
- Keep the power cord away from hot surfaces.
- If you notice any unexplained changes in the performance of the device, if it is making unusual sounds, if the device or the power supply are dropped or mishandled, or if the enclosure is broken, discontinue use and contact your care provider or your ResMed Service Center.
- Do not open or modify the device. There are no user serviceable parts inside. Repairs and servicing should only be performed by an authorized ResMed service agent.
- Beware of electrocution. Do not immerse the device, power supply or power cord in water. If liquids are spilled into or onto the device, unplug the device and let the parts dry. Always unplug the device before cleaning and make sure that all parts are dry before plugging it back in.
- Supplemental oxygen must not be used while smoking or in the presence of an open flame.
- Always make sure that the device is turned on and airflow generated before the oxygen supply is turned on. Always turn the oxygen supply off before the device is turned off, so that unused oxygen does not accumulate within the device enclosure and create a risk of fire.
- Do not perform any maintenance tasks while the device is in operation.
- The device should not be used adjacent to or stacked with other equipment. If adjacent or stacked use is necessary, the device should be observed to verify normal operation in the configuration in which it will be used.
- The use of accessories other than those specified for the device is not recommended. They may result in increased emissions or decreased immunity of the device.
- Regularly check the antibacterial filter for signs of moisture or other contaminants, particularly during nebulization or humidification. Failure to do so could result in increased breathing system resistance.


## CAUTION

- Use only ResMed parts and accessories with the device. Non-ResMed parts may reduce the effectiveness of the treatment and/or damage the device.
- Use only vented masks recommended by ResMed or by the prescribing doctor with this device. Fitting the mask without the device blowing air can result in rebreathing of exhaled air. Make sure that the mask vent holes are kept clear and unblocked to maintain the flow of the fresh air into the mask.
- Be careful not to place the device where it can be bumped or where someone is likely to trip over the power cord.
- Blocking the air tubing and/or air inlet of the device while in operation could lead to overheating of the device.
- Keep the area around the device dry, clean and clear of anything (eg, clothes or bedding) that could block the air inlet or cover the power supply unit.
- Do not place the device on its side as water might get into the device.
- Incorrect system setup may result in incorrect mask pressure reading. Ensure the system is correctly set up.
- Do not use bleach, chlorine, alcohol, or aromatic-based solutions, moisturizing or antibacterial soaps or scented oils to clean the device, the water tub or air tubing. These solutions may cause damage or affect the humidifier performance and reduce the life of the products.



- If you use the humidifier, always place the device on a level surface lower than your head to prevent the mask and air tubing from filling with water.
- Leave the water tub to cool for ten minutes before handling to allow the water to cool and to make sure that the water tub is not too hot to touch.
- Make sure that the water tub is empty before transporting the device.

## Technical specifications

|   |  |
|---|--|
| <b>90W power supply unit</b>  |  |
| AC input range:   | 100–240V, 50–60Hz 1.0–1.5A, Class II<br>115V, 400Hz 1.5A, Class II (nominal for aircraft use)  |
| DC output:  | 24V  3.75A  |
| Typical power consumption:  | 53W (57VA)   |
| Peak power consumption:   | 104W (108VA)   |
| <b>Environmental conditions</b>   |  |
| Operating temperature:  | +41°F to +95°F (+5°C to +35°C)<br><b>Note:</b> The air flow for breathing produced by this therapy device can be higher than the temperature of the room. Under extreme ambient temperature conditions (104°F/40°C) the device remains safe. |
| Operating humidity:   | 10 to 95% relative humidity, non-condensing  |
| Operating altitude:   | Sea level to 8,500' (2,591 m); air pressure range 1013 hPa to 738 hPa  |
| Storage and transport temperature:  | -4°F to +140°F (-20°C to +60°C)  |
| Storage and transport humidity:   | 5 to 95% relative humidity, non-condensing   |
| <b>Electromagnetic compatibility</b>  |  |
| The AirSense 10 complies with all applicable electromagnetic compatibility requirements (EMC) according to IEC 60601-1-2:2014, for residential, commercial and light industry environments. It is recommended that mobile communication devices are kept at least 1 m away from the device. |  |
| Information regarding the electromagnetic emissions and immunity of this ResMed device can be found on <a href="http://www.resmed.com/downloads/devices">www.resmed.com/downloads/devices</a>   |  |
| <b>Classification: IEC 60601-1:2005+A1:2012</b>   |  |
| Class II (double insulation), Type BF, Ingress protection IP22.   |  |
| <b>Sensors</b>  |  |
| Pressure sensor:  | Internally located at device outlet, analog gauge pressure type, -5 to +45 cm H <sub>2</sub> O   |
| Flow sensor:  | Internally located at device inlet, digital mass flow type, -70 to +180 L/min  |
| <b>Maximum single fault steady pressure</b>   |  |
| Device will shut down in the presence of a single fault if the steady state pressure exceeds:   |  |
| 30 cm H <sub>2</sub> O for more than 6 sec or 40 cm H <sub>2</sub> O for more than 1 sec.   |  |

---

## Sound

Pressure level measured according to ISO 80601-2-70:2015 (CPAP mode):

SlimLine: 25 dBA with uncertainty of 2 dBA

Standard: 25 dBA with uncertainty of 2 dBA

SlimLine or Standard and humidification: 27 dBA with uncertainty of 2 dBA

Power level measured according to ISO 80601-2-70:2015 (CPAP mode):

SlimLine: 33 dBA with uncertainty of 2 dBA

Standard: 33 dBA with uncertainty of 2 dBA

SlimLine or Standard and humidification: 35 dBA with uncertainty of 2 dBA

Declared dual-number noise emission values in accordance with ISO 4871:1996.

---

## Physical - device and water tub

Dimensions (H x W x D): 4.57" x 10.04" x 5.91"  
(116 mm x 255 mm x 150 mm)

Air outlet (complies with ISO 5356-1:2004): 22 mm

Weight (device and standard water tub): 44 oz (1248 g)

Weight (device and cleanable water tub): 44 oz (1248 g)

Housing construction: Flame retardant engineering thermoplastic

Water capacity: To maximum fill line 380 mL

Standard water tub - material: Injection molded plastic, stainless steel and silicone seal

Cleanable water tub - material: Injection molded plastic, stainless steel and silicone seal

---

## Temperature

Maximum heater plate: 154°F (68°C)

Cut-out: 165°F (74°C)

Maximum gas temperature: ≤ 106°F (≤ 41°C)

---

## Air filter

Standard: Material: Polyester non woven fiber  
Average arrestance: >75% for ~7 micron dust

Hypoallergenic: Material: Acrylic and polypropylene fibers in a polypropylene carrier  
Efficiency: >98% for ~7-8 micron dust; >80% for ~0.5 micron dust

---

## Aircraft use

ResMed confirms that device meets the Federal Aviation Administration (FAA) requirements (RTCA/DO-160, section 21, category M) for all phases of air travel.

---

## Wireless module

Technology used: 4G (USA and Canada only)  
3G (USA and Canada only)  
2G GSM (all regions except USA and Canada)

---

FCC ID: 2ACHL-AIR104G, 2ACHL-AIR103G

IC: 9103A-AIR104G, 9103A-AIR103G

The AirSense 10 device complies with FCC Rules and Industry Canada rules.

The AirSense 10 device should be used at a minimum distance of 0.8" (2 cm) from the body during operation.

Additional information regarding the FCC Rules and IC compliance for this device can be found on [www.resmed.com/downloads/devices](http://www.resmed.com/downloads/devices)

---

## Operating pressure range


AutoSet, AutoSet For Her, Elite, CPAP 4 to 20 cm H<sub>2</sub>O

---

## Supplemental oxygen

Maximum flow: 4 L/min

### Pneumatic flow path



1. Flow sensor
2. Blower
3. Pressure sensor
4. Mask
5. Air tubing
6. Water tub
7. Device
8. Inlet filter

### Design life

Device, power supply unit: 5 years  
Cleanable water tub: 2.5 years  
Standard water tub, air tubing: 6 months

### General

The patient is an intended operator.

### Humidifier performance

| Mask Pressure<br>cm H <sub>2</sub> O | RH output % at<br>63°F (17°C) ambient<br>temperature | RH output % at<br>72°F (22°C) ambient<br>temperature | Nominal system output AH <sup>1</sup> , BTPS <sup>2</sup> |           |
|--------------------------------------|--|--|---|-----------|
|                                      | Setting 4  | Setting 8  | Setting 4   | Setting 8 |
| 4                                    | 85   | 100  | 6   | >10       |
| 10                                   | 85   | 100  | 6   | >10       |
| 20                                   | 85   | 90   | 6   | >10       |

<sup>1</sup> AH - Absolute Humidity in mg/L

<sup>2</sup> BTPS - Body Temperature Pressure Saturated

### Air tubing

| Air tubing         | Material                                   | Length       | Inner diameter |
|--------------------|--|--------------|----------------|
| ClimateLineAir     | Flexible plastic and electrical components | 6'6" (2 m)   | 0.6" (15 mm)   |
| ClimateLineAir Oxy | Flexible plastic and electrical components | 6'4" (1.9 m) | 0.75" (19 mm)  |
| SlimLine           | Flexible plastic                           | 6' (1.8 m)   | 0.6" (15 mm)   |
| Standard           | Flexible plastic                           | 6'6" (2 m)   | 0.75" (19 mm)  |

Heated air tubing temperature cut-out: ≤ 106°F (≤ 41°C)

### Notes:

- The manufacturer reserves the right to change these specifications without notice.
- The electrical connector end of the heated air tubing is only compatible with the air outlet at the device end and should not be fitted to the mask.
- Do not use electrically conductive or antistatic air tubing.
- The temperature and relative humidity settings displayed are not measured values.

## Displayed values

| Value                          | Range                    | Display resolution      |
|--------------------------------|--------------------------|-------------------------|
| Pressure sensor at air outlet: |                          |                         |
| Mask pressure                  | 4–20 cm H <sub>2</sub> O | 0.1 cm H <sub>2</sub> O |
| Flow derived values:           |                          |                         |
| Leak                           | 0–120 L/min              | 1 L/min                 |

| Value                                       | Accuracy  |
|---|---|
| Pressure measurement <sup>1</sup> :         |   |
| Mask pressure <sup>2</sup>                  | ±[0.5 cm H <sub>2</sub> O + 4% of measured value]                                 |
| Flow and flow derived values <sup>1</sup> : |   |
| Flow  | ±6 L/min or 10% of reading, whichever is greater, at 0 to 150 L/min positive flow |
| Leak <sup>2</sup>                           | ±12 L/min or 20% of reading, whichever is greater, 0 to 60 L/min                  |

<sup>1</sup> Results are expressed as STPD (Standard Temperature and Pressure, Dry).

<sup>2</sup> Accuracy may be reduced by the presence of leaks, supplemental oxygen, tidal volumes <100 mL or minute ventilation <3 L/min.

## Measurement system uncertainties

In accordance with ISO 80601-2-70:2015 the measurement uncertainty of the manufacturer's test equipment is:

|                                   |   |
|-----------------------------------|---|
| For measures of flow              | ± 1.5 L/min or ± 2.7% of reading (whichever is greater) |
| For measures of volume (< 100 mL) | ± 5 mL or 6% of reading (whichever is greater)          |
| For measures of volume (≥ 100 mL) | ± 20 mL or 3% of reading (whichever is greater)         |
| For measures of pressure          | ± 0.15 cm H <sub>2</sub> O                              |
| For measures of time              | ± 10 ms   |

## Pressure accuracy

### Maximum static pressure variation at 10 cm H<sub>2</sub>O according to ISO 80601-2-70:2015

|                        | Standard air tubing       | SlimLine air tubing       |
|------------------------|---------------------------|---------------------------|
| Without humidification | ± 0.5 cm H <sub>2</sub> O | ± 0.5 cm H <sub>2</sub> O |
| With humidification    | ± 0.5 cm H <sub>2</sub> O | ± 0.5 cm H <sub>2</sub> O |

### Maximum dynamic pressure variation according to ISO 80601-2-70:2015

Device without humidification and Standard air tubing / Device with humidification and Standard air tubing

| Pressure [cm H <sub>2</sub> O] | 10 BPM    | 15 BPM    | 20 BPM    |
|--------------------------------|-----------|-----------|-----------|
| 4                              | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |
| 8                              | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |
| 12                             | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |
| 16                             | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |
| 20                             | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |

Device without humidification and SlimLine air tubing / Device with humidification and SlimLine air tubing

| Pressure [cm H <sub>2</sub> O] | 10 BPM    | 15 BPM    | 20 BPM    |
|--------------------------------|-----------|-----------|-----------|
| 4                              | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |
| 8                              | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |
| 12                             | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |
| 16                             | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |
| 20                             | 0.5 / 0.5 | 0.5 / 0.5 | 0.8 / 0.8 |


























## Flow (maximum) at set pressures

The following are measured accordingly to ISO 80601-2-70:2015 at the end of the specified air tubing:

| Pressure            | AirSense 10 and Standard | AirSense 10, humidification and Standard | AirSense 10 and SlimLine | AirSense 10, humidification and ClimateLineAir |
|---------------------|--------------------------|--|--------------------------|--|
| cm H <sub>2</sub> O | L/min                    | L/min                                    | L/min                    | L/min  |
| 4                   | 180                      | 143                                      | 162                      | 151  |
| 8                   | 168                      | 135                                      | 151                      | 142  |
| 12                  | 157                      | 136                                      | 140                      | 135  |
| 16                  | 144                      | 134                                      | 128                      | 121  |
| 20                  | 131                      | 123                                      | 117                      | 109  |

## Symbols

The following symbols may appear on the product or packaging.

 Read instructions before use. 
  Indicates a warning or caution. 
  Follow instructions before use. 
  Manufacturer. 
  European Authorized Representative. 
  Batch code. 
  Catalog number. 
  Serial number. 
  Device number. 
  On / Off. 
  Device weight. 
  Protected against finger sized objects and against dripping water when tilted up to 15 degrees from specified orientation. 
  Direct current. 
  Type BF applied part. 
  Class II equipment. 
  Humidity limitation. 
  Temperature limitation. 
  Non-ionising radiation. 
  China pollution control logo 1. 
  China pollution control logo 2. 
 **Rx Only** Prescription only (In the US, Federal law restricts these devices to sale by or on the order of a physician). 
  Maximum water level. 
  Use distilled water only. 
  Operating altitude. 
  Atmospheric pressure limitation. 
  Complies with RTCA DO-160 section 21, category M.



## Environmental information

This device should be disposed of separately, not as unsorted municipal waste. To dispose of your device, you should use appropriate collection, reuse and recycling systems available in your region. The use of these collection, reuse and recycling systems is designed to reduce pressure on natural resources and prevent hazardous substances from damaging the environment.

If you need information on these disposal systems, please contact your local waste administration. The crossed-bin symbol invites you to use these disposal systems. If you require information on collection and disposal of your ResMed device please contact your ResMed office, local distributor or go to [www.resmed.com/environment](http://www.resmed.com/environment).

## Servicing

The AirSense 10 device is intended to provide safe and reliable operation when operated in accordance with the instructions provided by ResMed. ResMed recommends that the AirSense 10 device be inspected and serviced by an authorized ResMed Service Centre if there is any sign of wear or concern with device function. Otherwise, service and inspection of the products generally should not be required during their design life.

## Limited warranty

ResMed Ltd (hereafter 'ResMed') warrants that your ResMed product shall be free from defects in material and workmanship from the date of purchase for the period specified below.

| Product   | Warranty period |
|---|-----------------|
| <ul style="list-style-type: none"><li>• Mask systems (including mask frame, cushion, headgear and tubing)—excluding single-use devices</li><li>• Accessories—excluding single-use devices</li><li>• Flex-type finger pulse sensors</li><li>• Humidifier water tubs</li></ul>    | 90 days         |
| <ul style="list-style-type: none"><li>• Batteries for use in ResMed internal and external battery systems</li></ul>   | 6 months        |
| <ul style="list-style-type: none"><li>• Clip-type finger pulse sensors</li><li>• CPAP and bilevel device data modules</li><li>• Oximeters and CPAP and bilevel device oximeter adapters</li><li>• Humidifier cleanable water tubs</li><li>• Titration control devices</li></ul> | 1 year          |
| <ul style="list-style-type: none"><li>• CPAP, bilevel and ventilation devices (including external power supply units)</li><li>• Humidifiers</li><li>• Battery accessories</li><li>• Portable diagnostic/screening devices</li></ul>   | 2 years         |

This warranty is only available to the initial consumer. It is not transferable.

If the product fails under conditions of normal use, ResMed will repair or replace, at its option, the defective product or any of its components.

This Limited Warranty does not cover: a) any damage caused as a result of improper use, abuse, modification or alteration of the product; b) repairs carried out by any service organization that has not been expressly authorized by ResMed to perform such repairs; and c) any damage or contamination due to cigarette, pipe, cigar or other smoke.

Warranty is void on product sold, or resold, outside the region of original purchase.

Warranty claims on defective product must be made by the initial consumer at the point of purchase.

This warranty replaces all other expressed or implied warranties, including any implied warranty of merchantability or fitness for a particular purpose. Some regions or states do not allow limitations on how long an implied warranty lasts, so the above limitation may not apply to you.

ResMed shall not be responsible for any incidental or consequential damages claimed to have resulted from the sale, installation or use of any ResMed product. Some regions or states do not allow the exclusion or limitation of incidental or consequential damages, so the above limitation may not apply to you.

This warranty gives you specific legal rights, and you may also have other rights which vary from region to region. For further information on your warranty rights, contact your local ResMed dealer or ResMed office.

## Further information

If you have any questions or require additional information on how to use the device, contact your care provider.

**ResMed Ltd**

1 Elizabeth Macarthur Drive  
Bella Vista NSW 2153 Australia

**DISTRIBUTED BY**

ResMed Corp 9001 Spectrum Center Boulevard San Diego CA 92123 USA  
**EC REP** ResMed (UK) Ltd 96 Jubilee Ave Milton Park Abingdon Oxfordshire OX14 4RW UK

See ResMed.com for other ResMed locations worldwide. Air10, AirSense, AutoSet, ClimateLine, HumidAir and SlimLine are trademarks and/or registered trademarks of the ResMed family of companies. For patent and other intellectual property information, see ResMed.com/ip. SD Logo is a trademark of SD-3C, LLC. © 2016 ResMed Ltd. 378720/1 2016-12

ResMed.com



378720