



Patient produces moisture & heat upon exhalation

ActiveAir vent

Heat and moisture retained in HumidX

Must be replaced every 30 days for optimal performance

Co₂ washed out through ActiveAir vent