6 Quick Steps to Recharge Your Battery



INOGENONE GALLERY Charger Guide

Your Equipment Provider:

Your Physician:



PMS 661 main color/651 swish

301 Coromar Drive, Goleta, CA 93117 1-877-4-INOGEN (Toll Free) 1-877-446-6436 (Toll Free) 1-805-562-0515 (International) info@inogen.net www.inogen.com

96-06732-00-01 C





Plug the External Battery Charger AC power supply cord into an electrical outlet.

INOGENONE GA External Battery Charger

Your Inogen One G4 External Battery Charger (BA-403) will charge your Inogen One G4 single and double batteries. It takes approximately 2 hours to charge a fully depleted single battery, and approximately 4 hours to charge a fully depleted double battery. Note: the battery will also recharge in the Inogen One G4 when powered by the AC Power Supply, or DC Power Cable.



Plug the External Battery Charger AC power supply into the battery charger.



Slide the charger onto the Inogen One G4 Battery by clicking and locking into the charger.



BATTERY CHARGER
Inogen®

1000

When the battery is in the correct position, a solid red light will indicate that the battery is charging.



Press battery latch down and slide charger off battery.



When the green light illuminates, the battery is fully charged*.



©2021 Inogen. All rights reserved www.inogen.com

*If the red light is flashing, unplug the unit and complete steps 2-3 again. If a red flashing light continues after these steps, contact your Equipment Provider.